



CERTIFICATE OF ANALYSIS

Alkaline Phosphatase Conjugate Stabilizer, 1X				
Catalog#	Size	Lot#	Expiry at 2-8°C	
3323	Bulk	21F52	6/17/2023	
6270	25 mL	NA	NA	
6271	100 mL	NA	NA	
6273	1L	NA	NA	
6274	10 L	21F53	6/17/2023	

Specifications	Criteria	Lot Data	Results
рН	6.0 ± 0.50	6.0	Pass
Accelerated Stability Test: AP antibody conjugate is diluted to working concentration, stored at 37°C for 1 week, and then tested for retained activity.	≥85% (relative to 4°C control)	Conforms	Pass
Appearance	Clear, light yellow liquid	Conforms	Pass
Storage	Refrigerated	Refrigerated	Pass

This is an assurance that the specified lot conforms to all requirements of the product specifications. This information is believed to be accurate and is offered in good faith.

Quality Assurance	Trans Myphs		
	Tracy Murphy		
Date	2/16/2021		
	2/16/2021		

9401 James Avenue S., #155 Bloomington, MN 55431 USA TOLL-FREE 1-800-829-3194 LOCAL 952-888-8788